Iron Low Range (Fe)

Method - Colorimetric

After adding the sample into the measurement cell, some reagents are added in order to adjust the solution to the desired conditions (pH, valence's elements, etc.). Then, a blank is done to correct any temperature or turbidity disturbance. Subsequently, a last reagent is added, and it reacts with solution developing a color, which is measured using a correct wavelength. Thanks to the photometer used, the result achieves a great accuracy

Principle of measurement

Any ferric iron is reduced to the ferrous state by means of the hydroxylamine hydrochloride. The ferrous ions react with the Ferrozine to form a pink complex when the buffer brings the solution within the range 5 to 7 pH.

Advantages of the method

The method enables iron to be measured at very low concentrations. The reagents are stable. If total iron has to be measured the stirring time may be increased to ensure that the iron has been completely dissolved by the acidic medium.

Specifications

Range: From 0 to 100 ppb / 500 ppb.
Adjustable higher concentrations with internal dilution.
Accuracy: ±2%
Repeatability: ±2%
Resolution: 0.1 ppb
Analysis time: around 15 minutes
Calibration: two point
LED Wavelength: 545 nm

Applications

• Potable Water Treatment Plants (PWTP)
And more...

Reagents consumption

• Reagent 1: 1 ml / analysis — 0.75L / month
• Reagent 2: 2 ml / analysis — 1.50L / month
Monthly consumption calculated assuming 1 analysis per hour.

Visit our web

Instru | INSTRUMENTACIÓN ANALÍTICA Instran®

Solutions
for those
moving the world

www.instran.net

Instru office
c/ Penedés, 46
08820 El Prat de Llobregat
Barcelona • Tel.+34 934 787 161
instran@instran.net
c.salinas@instran.net

Instran®

On-line Analyzer
Water quality parameters

Iron
Low Range
(Fe)

We offer you all the solutions you need to keep
moving the world forward